

## Diagnostic reference levels in intraoral dental radiography in Korea

Eun-Kyung Kim, Won-Jeong Han, Jin-Woo Choi, Yun-Hoa Jung\*, Suk-Ja Yoon\*\*, Jae-Seo Lee\*\*

Department of Oral and Maxillofacial Radiology, Dankook University College of Dentistry, Cheonan, Korea

\*Department of Oral and Maxillofacial Radiology, College of Dentistry, Pusan National University, Yangsan, Korea

\*\*Department of Oral and Maxillofacial Radiology, School of Dentistry, Chonnam National University, Gwangju, Korea

### ABSTRACT

**Purpose:** The objectives of this study were to survey the radiographic exposure parameters, to measure the patient doses for intraoral dental radiography nationwide, and thus to establish the diagnostic reference levels (DRLs) in intraoral dental X-ray examination in Korea.

**Materials and Methods:** One hundred two intraoral dental radiographic machines from all regions of South Korea were selected for this study. Radiographic exposure parameters, size of hospital, type of image receptor system, installation duration of machine, and type of dental X-ray machine were documented. Patient entrance doses (PED) and dose-area products (DAP) were measured three times at the end of the exit cone of the X-ray unit with a DAP meter (DIAMENTOR M4-KDK, PTW, Freiburg, Germany) for adult mandibular molar intraoral dental radiography, and corrections were made for room temperature and pressure. Measured PED and DAP were averaged and compared according to the size of hospital, type of image receptor system, installation duration, and type of dental X-ray machine.

**Results:** The mean exposure parameters were 62.6 kVp, 7.9 mA, and 0.5 second for adult mandibular molar intraoral dental radiography. The mean patient dose was 2.11 mGy (PED) and 59.4 mGycm<sup>2</sup> (DAP) and the third quartile one 3.07 mGy (PED) and 87.4 mGycm<sup>2</sup> (DAP). Doses at university dental hospitals were lower than those at dental clinics ( $p < 0.05$ ). Doses of digital radiography (DR) type were lower than those of film-based type ( $p < 0.05$ ).

**Conclusion:** We recommend 3.1 mGy (PED), 87.4 mGycm<sup>2</sup> (DAP) as the DRLs in adult mandibular molar intraoral dental radiography in Korea. (*Imaging Sci Dent* 2012; 42 : 237-42)

**KEY WORDS:** Radiation Protection; Radiation Dosage; Radiography, Dental

### Introduction

Recently, patient exposure to medical and dental X-ray examination has grown rapidly and diagnostic radiology represents the largest source of artificial radiation which is comparable to natural background exposure.<sup>1</sup> For patient protection, the principles of justification and optimization should be followed. All radiographic examinations have to show a potential benefit to the patient weighing against

the potential risk. After they are justified, the radiographic exposure should be kept as low as reasonably achievable, taking into account economic and societal factors. The objective of this optimization is to decrease the total patient dose of radiation without compromising diagnosis. However, the optimization process is a complicated procedure. Many international and national surveys have shown a wide distribution of patient doses for the same type of radiographic examination.<sup>2</sup> The concept of diagnostic reference levels (DRLs) has been introduced and applied to different radiodiagnostic examinations in the medical and dental field.<sup>2</sup> DRLs are dose levels in medical radiodiagnostic practices for typical examinations for groups of standard-sized patients or standard phantoms for broadly defined types of equipment.<sup>3-5</sup> These are based on the third quar-

\*This research was supported by a grant (09142KFDA510) from Korea Food and Drug Administration in 2009.

Received May 30, 2012; Revised June 22, 2012; Accepted July 7, 2012

Correspondence to : Prof. Eun-Kyung Kim

Department of Oral and Maxillofacial Radiology, Dankook University College of Dentistry, 119 Dandae-ro, Dongnam-gu, Cheonan, Chungnam 330-714, Korea  
Tel) 82-41-550-1924, Fax) 82-41-556-7127, E-mail) ekkim@dankook.ac.kr

**Table 1.** Regional distribution of dental radiographic equipment selected for this study

Seoul	Gyeonggi	Chungnam	Chungbuk	Busan	Gyeongnam	Daegu	Gwangju	Jeonnam	Jeonbuk	Jeju	Total
24	11	16	6	26	5	4	17	10	4	3	126
(24)	(11)	(16)	(6)	(26)	(5)	(4)	(4)	(6)	(0)	(0)	(102)

The number without parenthesis indicates that of dental radiographic equipment used for the survey of radiographic exposure parameters. The number within parentheses indicates that of dental radiographic equipment used for the measurement of patient dose.

tile values for the distributions of doses found in the national or regional surveys, that is, 75% of hospitals are giving patient doses below these values. The 25% of hospitals above these are most urgently in need of better quality control, for example, identification of inadequate techniques or machine malfunctions.<sup>6</sup>

Dose quantities adopted for DRLs are patient entrance dose (PED), entrance surface dose (ESD), dose area product (DAP), and other dose-related quantities. PED is defined as the absorbed dose (to air) measured at the end of the spacer 'cone' for a typical examination (adult mandibular molar) without backscatter from the patient.<sup>6,7</sup> ESD is defined as the absorbed dose to air at the point of intersection of the X-ray beam axis with the entrance surface of the patient, including backscattered radiation from the patient.<sup>8</sup> DAP is defined as the absorbed dose to air averaged over the area of the X-ray beam in a plane perpendicular to the beam axis, multiplied by the area of the beam in the same plane, namely the integral of the dose across the X-ray beam.<sup>8</sup> This is conveniently measured with special large-area ionization chambers (DAP meters), which intercept the entire cross section of the beam.<sup>8</sup> DAP correlates reasonably well with radiation risk, as the number of interactions within the patient is proportional to both dose and field size.<sup>9</sup> In dental radiology, PED was recommended and commonly used for the setting of DRLs, and it differs from the quantity ESD commonly used in general medical radiography by not including radiation backscattered from the patient.<sup>6,7</sup> Recently, DAP has been recommended for the setting of DRLs in intraoral,<sup>8</sup> panoramic,<sup>10,11</sup> cephalometric,<sup>12</sup> and cone-beam CT examinations.<sup>13</sup>

The objectives of this study were to survey the radiographic exposure parameters, and measure the patient dose for intraoral dental radiography nationwide, and thus to establish the DRLs in intraoral dental X-ray examination in South Korea. This was the first nationwide investigation for the development of DRLs in intraoral dental radiography. The patient doses were measured with both dose quantities, PED and DAP, in order to compare them with previous reports from other countries.



**Fig. 1.** The ionization chamber of DAP meter is positioned at the end of the exit cone of the intraoral X-ray machine for PED and DAP measurement.

## Materials and Methods

One hundred twenty-six intraoral dental radiographic machines in 95 dental institutions were selected from all regions of South Korea for this study. The radiographic exposure parameters for intraoral dental radiographic examination were surveyed with 126 intraoral dental radiographic machines. The patient doses were measured with a DAP meter (Diamentor M4-KDK, PTW, Freiburg, Germany) for 102 intraoral dental radiographic machines from March to October 2009 (Table 1). Due to the mechanical failure of the ionization chamber of the DAP meter, the last 24 radiographic machines were excluded from the patient dose measurement.

Radiographic exposure parameters (kV, mA, exposure time, focal spot-skin distance), size of hospital (university dental hospitals, dental hospital, dental clinic, public health center), type of image receptor system (film-based type, digital radiography type, computed radiography type), installation duration of machines (5 years or less, 6 years or more), and type of dental X-ray machine (wall-mounted fixed type, hand-held portable type) were documented.

PED and DAP were measured three times at the end of the exit cone of the X-ray unit with a DAP meter for adult mandibular molar intraoral dental radiography (Fig. 1), and corrections were made for room temperature and pressure. The PED was expressed in terms of mGy and DAP in mGycm<sup>2</sup>. The measured PED and DAP were averaged and compared according to the size of hospital, type of image receptor system, installation duration, and type of dental X-ray machine. Independent t-tests and ANOVA tests were performed for the comparisons using SPSS 12.0.1 for Windows (SPSS Inc., Chicago, IL, USA).

**Table 2.** Exposure parameters for adult mandibular molar intraoral dental radiography

	Tube voltage (kV)	Tube current (mA)	Exposure time (s)	FSD (mm)
Minimum	60	1	0.02	105
Median	60	10	0.5	205
Maximum	70	15	2	344
Mean	62.6	7.9	0.5	218

FSD: focal spot - skin distance measured as the distance from focal spot to the end of the exit cone of the intraoral x-ray machine

## Results

The ranges of exposure parameters for adult mandibular molar intraoral dental radiography were 60-70 kV, 1-15 mA, and 0.02-2 second and their means were 62.6 kV, 7.9 mA, and 0.5 seconds. The range of focal spot-skin distance (FSD) was 105-344 mm with a mean of 218 mm (Table 2). In comparison of radiographic exposure parameters according to the size of hospital, the university dental hospitals showed the highest tube voltage and the shortest exposure time (Table 3). In comparison according to the type of image receptor system, the digital radiography (DR) type showed a higher tube voltage, lower tube current, and shorter exposure time than the film-based type (Table 4).

The mean patient doses were 2.11 mGy (PED) and 59.4 mGycm<sup>2</sup> (DAP), and the third quartile ones were 3.07 mGy (PED) and 87.4 mGycm<sup>2</sup> (DAP) (Table 5, Figs. 2 and 3). The mean patient doses at the university dental hospitals were lower than those at the dental clinics ( $p < 0.05$ ) (Table 6). The mean patient doses of the DR type were lower than those of the film-based type ( $p < 0.05$ ) (Table 7). The mean

**Table 3.** Comparison of radiographic exposure parameters according to the size of hospital


	No. of machines	%	Tube voltage (kV)	Tube current (mA)	Exposure time (s)	FSD (mm)
University dental hospital	32	25	65.8 <sup>*,†,‡</sup>	7.6	0.2 <sup>*,†</sup>	242 <sup>*,†</sup>
Dental hospital	12	9	62.5 <sup>*</sup>	7.2	0.5	193 <sup>*</sup>
Dental clinic	75	60	61.5 <sup>†</sup>	8.1	0.7 <sup>*</sup>	208 <sup>†</sup>
Public health center	7	6	60.0 <sup>‡</sup>	8.9	0.6 <sup>†</sup>	251

<sup>\*,†,‡</sup>: statistically significant at  $p < 0.05$


**Table 4.** Comparison of radiographic exposure parameters according to the type of image receptor system

	No. of machines	%	Tube voltage (kV)	Tube current (mA)	Exposure time (s)	FSD (mm)
Film-based type	51	41	61.3 <sup>*</sup>	9.4 <sup>†</sup>	0.8 <sup>†</sup>	231
DR type	71	56	63.6 <sup>*</sup>	6.9 <sup>†</sup>	0.4 <sup>†</sup>	209
CR type	4	3	62.5	7.3	0.6	199

<sup>\*,†,‡</sup>: statistically significant at  $p < 0.05$ . DR: Digital Radiography, CR: Computed Radiography



**Fig. 2.** Measured PED values for intraoral dental radiography.



**Fig. 3.** Measured DAP values for intraoral dental radiography.

**Table 5.** Patient doses for adult mandibular molar intraoral dental radiography

	PED (mGy)	DAP (mGycm <sup>2</sup> )
Minimum	0.21	5.5
Maximum	10.98	304.2
Mean	2.11	59.4
3rd quartile	3.07	87.4

**Table 6.** Comparison of mean patient doses according to the size of hospital

	No. of machines	%	PED (mGy)	DAP (mGycm <sup>2</sup> )
University dental hospital	23	22	0.91*	25.1 <sup>†</sup>
Dental hospital	12	12	2.17	60.2
Dental clinic	61	60	2.54*	71.6 <sup>†</sup>
Public health center	6	6	2.21	64.7

\*.†: statistically significant at  $p < 0.05$

**Table 7.** Comparison of mean patient dose according to the type of image receptor system

Type of image receptor	No. of machines	%	PED (mGy)	DAP (mGycm <sup>2</sup> )
Film-based type	43	42	3.05*	84.4 <sup>†</sup>
DR type	56	55	1.35*	38.8 <sup>†</sup>
CR type	3	3	2.80	83.4

\*.†: statistically significant at  $p < 0.05$ . DR: Digital Radiography, CR: Computed Radiography

patient doses did not show statistically significant differences according to equipment installation duration and type of dental X-ray system (Tables 8 and 9).

### Discussion

National surveys on medical and dental radiography have been performed in many countries. Among them, the

**Table 8.** Comparison of mean patient doses according to the installation duration

Installation duration	No. of machines	%	PED (mGy)	DAP (mGycm <sup>2</sup> )
5 years or less	50	49	1.70	48.6
6 years or more	52	51	2.50	69.7

**Table 9.** Comparison of mean patient dose according to type of dental X-ray machine

Type of dental X-ray machine	No. of machines	%	PED (mGy)	DAP (mGycm <sup>2</sup> )
Wall-mounted	87	85	2.14	60.7
Hand-held	15	15	1.90	50.6

UK has reported national DRLs consistently through a series of five-yearly reviews of the National Patient Dose Database maintained by the Radiation Protection Division (RPD) of the Health Protection Agency (HPA).<sup>14</sup> In the 2005 review, they reported that the reference doses were on average about 16% lower than the corresponding values in the previous (2000) review, and were typically less than half the values of the original UK national reference doses that were derived from a survey in the mid-1980s.<sup>14</sup> The DRL for an adult mandibular molar intraoral radiograph recommended by the NRPB (the organization prior to the HPA in the UK) was 4 mGy in 1999<sup>6,14</sup> and it had fallen to 2.3 mGy at the next review in 2005.<sup>14</sup> According to the IAEA Basic Safety Standards (1996),<sup>15</sup> the guidance level of dose for periapical radiography was 7 mGy (ESD). Pope et al<sup>5</sup> carried out a comprehensive study in order to propose DRLs for intraoral radiology in Germany and reported the DAP value for mandibular molar radiographs to be 41.2 mGycm<sup>2</sup>. Tierris et al<sup>11</sup> reported 62 mGycm<sup>2</sup> (mean DAP) at 60 kV in Greece.

The 3rd quartile and mean values of the patient dose in the present study were 3.07 mGy and 2.11 mGy, respec-

tively, in dose quantity of PED, and 87.4 mGycm<sup>2</sup> and 59.4 mGycm<sup>2</sup>, respectively, in dose quantity of DAP. From the measured patient dose values, we could observe a wide range in patient doses among different dental facilities, as has been reported in other countries.<sup>5,7</sup> In our study, the range was from 0.21 to 10.98 mGy, with a factor of around 50. The HPA report in the UK's 2005 review showed a much larger range from 0.02 to 30 mGy with a factor of 1500 between the lowest and the highest doses. It was reported that around 15% of dentists were using digital systems for intraoral radiography at that time in the UK.<sup>14</sup> The 102 intraoral dental radiographic machines measured in this study comprised 43 of the film-based type, 56 DR type, and 3 CR type. Namely, about 58% of dentists used digital systems, particularly the DR type in Korea. The mean PED and DAP of the film-based type were 3.05 mGy and 84.4 mGycm<sup>2</sup>, respectively, and those of the DR type 1.35 mGy and 38.8 mGycm<sup>2</sup>, respectively, which were much lower than the values for the film-based type. The reason why the mean patient dose in university dental hospitals was the lowest was concluded to be because they used radiographic equipment with a high tube voltage and sensitive DR sensor system. Film-based systems generally showed high patient doses. Although the mean patient dose of the DR systems was lower than that of the film-based systems, the DR systems showed a wide variation in doses among different models, from low to high patient doses (including some even higher than those of the film type). Therefore, when selecting an intraoral DR sensor system, a dentist should consider the patient dose as well as the image quality for the optimization of intraoral dental radiography.

In the comparison according to equipment installation duration, the machines 5 years or less showed a lower mean patient dose than those 6 years or more; however, the difference was not statistically significant. This was assumed to be because some newer DR systems showed a considerably high patient dose.

Fifteen hand-held, portable intraoral dental radiographic machines (15%) were involved in this study. In the comparison according to type of dental X-ray system, the mean patient dose of hand-held systems was slightly lower than that of the wall-mount fixed systems, but the difference was not statistically significant. This was believed to be because some hand-held systems showed a much higher patient dose than the wall-mount fixed systems. The use of hand-held dental X-ray systems for general dental radiography remains in dispute because its use requires the operator's hand holding the X-ray tube housing. In this nation-

wide survey, it was observed that most of them were combined with a DR sensor, but a few systems were combined with X-ray film. It is recommended that dental practitioners do not use a hand-held dental X-ray system if possible. In case its use is needed, they should select a hand-held dental X-ray system with a low patient dose and use it in combination with a sensitive DR sensor. This was the first nationwide survey for the development of DRLs for intraoral dental radiography in South Korea. Consistently reported DRLs will assist in the ongoing reduction of patient radiation doses.

In conclusion, we recommend 3.1 mGy (PED) and 87.4 mGycm<sup>2</sup> (DAP) as the DRLs in adult mandibular molar intraoral dental radiography in South Korea based on this nationwide survey.

## References

1. White SC, Pharoah MJ. Oral Radiology: principles and interpretation. 6th ed. St. Louis: Mosby; 2008. p. 32-43.
2. Leitz W, Gron P, Servomaa A, Einarsson G, Olerud H. Nordic working group for medical x-ray diagnostics: diagnostic reference levels within x-ray diagnostics - experiences in the Nordic countries. In: Paile W, Editor. Radiation protection in the 2000s - Theory and practice. Nordic society for radiation protection: Proceedings of the XIII ordinary meeting. Turku/Åbo, Finland August 25-29, 2002. Helsinki: Radiation and Nuclear Safety Authority; 2003. p. 256-61.
3. Council directive 97/43/Euratom of 30 June 1997 on health protection of individuals against the dangers of ionizing radiation in relation to medical exposures, and repealing Directive 84/466/Euratom [Internet]. Luxembourg: European Commission; 1997 [cited 2012 April 30], Available from [http://ec.europa.eu/energy/nuclear/radioprotection/doc/legislation/9743\\_en.pdf](http://ec.europa.eu/energy/nuclear/radioprotection/doc/legislation/9743_en.pdf).
4. Diagnostic reference levels in medical imaging: review and additional advice. Ann ICRP 2001; 31 : 33-52.
5. Poppe B, Looe HK, Pfaffenberger A, Eenboom F, Chofor N, Sering M, et al. Radiation exposure and dose evaluation in intraoral dental radiology. Radiat Prot Dosimetry 2007; 123 : 262-7.
6. Napier ID. Reference doses for dental radiography. Br Dent J 1999; 186 : 392-6.
7. Gulson AD, Knapp TA, Ramsden PG. HPA-RPD-022. Doses to patients arising from dental X-ray examinations in the UK, 2002-2004. A review of dental X-ray protection service data [Internet]. Chilton: HPARPD; 2007 [cited 2012 April 30]. Available from [http://www.hpa.org.uk/web/HPAwebFile/HPAweb\\_C/1194947326586](http://www.hpa.org.uk/web/HPAwebFile/HPAweb_C/1194947326586).
8. Wall BF. Radiation protection dosimetry for diagnostic radiology patients. Radiat Prot Dosimetry 2004; 109 : 409-19.
9. Faulkner K, Broadhead DA, Harrison RM. Patient dosimetry measurement methods. Appl Radiat Isot 1999; 50 : 113-23.
10. Helmrot E, Alm Carlsson G. Measurement of radiation dose in dental radiology. Radiat Prot Dosimetry 2005; 114 : 168-71.
11. Tierris CE, Yakoumakis EN, Bramis GN, Georgiou E. Dose

- area product reference levels in dental panoramic radiology. *Radiat Prot Dosimetry* 2004; 111 : 283-7.
12. Looe HK, Eenboom F, Chofer N, Pfaffenberger A, Sering M, Rühmann A, et al. Dose-area product measurements and determination of conversion coefficients for the estimation of effective dose in dental lateral cephalometric radiology. *Radiat Prot Dosimetry* 2007; 124 : 181-6.
  13. SEDENTEXCT. Radiation protection: cone beam CT for dental and maxillofacial radiology. Evidence based guidelines 2011 (v2.0) [Internet]. SEDENTEXCT project; 2011 [cited 2012 April 30]. Available from: <http://www.sedentext.eu/content/guidelines-cbct-dental-and-maxillofacial-radiology>.
  14. Hart D, Hillier MC, Wall BF. HPA-RPD-029. Doses to patients from radiographic and fluoroscopic x-ray imaging procedures in the UK - 2005 review [Internet]. Chilton: HPARPD; 2007 [cited 2012 April 30]. Available from [http://www.hpa.org.uk/webc/HPAwebFile/HPAweb\\_C/1194947413167](http://www.hpa.org.uk/webc/HPAwebFile/HPAweb_C/1194947413167).
  15. IAEA. International basic safety standards for protection against ionizing radiation and for the safety of radiation source, jointly sponsored by FAO, IAEA, ILO, OECD/NEA, PAHO and WHO. Safety Series No. 115 [internet]. Vienna: IAEA; 1996 [cited 2012 April 30]. Available from [http://www-pub.iaea.org/mtcd/publications/pdf/ss-115-web/pub996\\_web-1a.pdf](http://www-pub.iaea.org/mtcd/publications/pdf/ss-115-web/pub996_web-1a.pdf).