

Clinical validity and precision of deep learning-based cone-beam computed tomography automatic landmarking algorithm

Jungeun Park¹, Seongwon Yoon^{2,3}, Hannah Kim^{3,4}, Youngjun Kim^{3,4},
Uilyong Lee^{5,*}, Hyungseog Yu^{6,*}

¹Department of Orthodontics, College of Dentistry, Yonsei University, Seoul, Korea

²College of Dentistry, Seoul National University, Seoul, Korea

³Imageworks Incorporated, Seoul, Korea

⁴Center for Bionics, Korea Institute of Science and Technology, Seoul, Korea

⁵Department of Oral and Maxillofacial Surgery, College of Dentistry, Chungang University Hospital, Seoul, Korea

⁶Department of Orthodontics, The Institute of Craniofacial Deformity, College of Dentistry, Yonsei University, Seoul, Korea

ABSTRACT

Purpose: This study was performed to assess the clinical validity and accuracy of a deep learning-based automatic landmarking algorithm for cone-beam computed tomography (CBCT). Three-dimensional (3D) CBCT head measurements obtained through manual and automatic landmarking were compared.

Materials and Methods: A total of 80 CBCT scans were divided into 3 groups: non-surgical (39 cases); surgical without hardware, namely surgical plates and mini-screws (9 cases); and surgical with hardware (32 cases). Each CBCT scan was analyzed to obtain 53 measurements, comprising 27 lengths, 21 angles, and 5 ratios, which were determined based on 65 landmarks identified using either a manual or a 3D automatic landmark detection method.

Results: In comparing measurement values derived from manual and artificial intelligence landmarking, 6 items displayed significant differences: R U6CP-L U6CP, R L3CP-L L3CP, S-N, Or_R-R U3CP, L1L to Me-GoL, and GoR-Gn/S-N ($P < 0.05$). Of the 3 groups, the surgical scans without hardware exhibited the lowest error, reflecting the smallest difference in measurements between human- and artificial intelligence-based landmarking. The time required to identify 65 landmarks was approximately 40-60 minutes per CBCT volume when done manually, compared to 10.9 seconds for the artificial intelligence method (PC specifications: GeForce 2080Ti, 64GB RAM, and an Intel i7 CPU at 3.6 GHz).

Conclusion: Measurements obtained with a deep learning-based CBCT automatic landmarking algorithm were similar in accuracy to values derived from manually determined points. By decreasing the time required to calculate these measurements, the efficiency of diagnosis and treatment may be improved. (*Imaging Sci Dent* 2024; 54: 240-50)

KEY WORDS: Cone-Beam Computed Tomography; Anatomic Landmarks; Cephalometry; Deep Learning; Orthognathic Surgery

This work was supported by the Korea Medical Device Development Fund grant, funded by the Korean government (Ministry of Science and ICT; Ministry of Trade, Industry and Energy; Ministry of Health & Welfare; Ministry of Food and Drug Safety) under Project No. RS-2020-KD000002.

Received January 22, 2024; Revised May 17, 2024; Accepted May 24, 2024

Published online August 12, 2024

*Correspondence to : Prof. Uilyong Lee

Department of Oral and Maxillofacial Surgery, College of Dentistry, Chungang University Hospital, 102, Heukseok-ro, Dongjak-gu, Seoul 06973, Korea
Tel) 82-02-6299-2875, E-mail) davidjoy@caumc.or.kr

Prof. Hyungseog Yu

Department of Orthodontics, The Institute of Craniofacial Deformity, College of Dentistry, Yonsei University, Seoul 03722, Korea
Tel) 82-02-2228-3104, E-mail) YUMICHAEL@yuhs.ac

Introduction

Cephalometric analysis is essential for diagnosis and treatment planning in orthodontic and orthognathic surgery. Traditionally, 2-dimensional (2D) cephalometric radiography has been used to evaluate the craniomaxillofacial (CMF) region. However, this imaging modality projects a 3-dimensional (3D) CMF structure onto a 2D plane, leading to image distortion. This distortion can manifest as the overlapping of anatomical structures and

Copyright © 2024 by Korean Academy of Oral and Maxillofacial Radiology

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/3.0>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

Imaging Science in Dentistry · pISSN 2233-7822 eISSN 2233-7830

the enlargement or reduction of specific areas.¹

Several experts have proposed the use of 3D cephalometric analysis with computed tomography or cone-beam computed tomography (CBCT) images.^{2,3} CBCT enables clinicians to visualize anatomy in 3 dimensions without overlap, providing comprehensive information on anatomical spatial relationships.⁴

Accurate landmarking is essential for proper diagnosis. However, manual landmarking can be repetitive and laborious, often yielding inconsistencies between and within practitioners. To alleviate the challenges associated with manual landmarking, numerous studies have explored the use of automatic landmark detection systems.^{5,6}

A wealth of information can be obtained from 3D cephalometry; however, 3D landmarking is difficult, labor-intensive, time-consuming, and heavily dependent on expertise and experience.^{7,8} The complexity of processing 3D data contributes to these challenges, as does the substantial effort required to create a 3D labeled dataset. Furthermore, since no open dataset is available for 3D CMF landmarks, many studies have had to rely on small datasets and have limited their measurements to landmarks on the bone surface.

As indicated above, research on 2D and 3D automatic landmarking is ongoing. However, few studies have examined the utility and clinical applicability of cephalometric analysis based on landmarks identified through automatic processes. The clinically acceptable margin of error for landmark placement depends on the error value when implemented in a clinical setting.⁹

When conducting cephalometric analyses, it is essential to ascertain the impact of errors in linear, angular, and ratio measurements across all 3 dimensions (x, y, and z coordinates).

In a prior study, Jeon and Lee compared 26 measurements obtained from a convolutional neural network (CNN)-based 2D automatic head landmarking system with those derived from conventional landmarking across 35 lateral head radiographs.¹⁰ Gupta et al. compared 51 measurements between a knowledge-based 3D automatic head landmarking system and manual landmarking for 30 CBCT scans.¹¹ Both studies concluded that automated cephalometric analysis is comparable in accuracy to manual calculations.^{10,11}

In dentistry, metal artifacts are commonly observed on CBCT images due to materials used in orthodontics, surgical applications, and dental restorations. The presence of metal artifacts along the path of the radiation beam causes photon depletion and scattering, resulting in char-

acteristic light and dark banding artifacts on the CBCT image.¹² These artifacts obscure the adjacent anatomy and impede diagnosis; furthermore, they can interfere with the image segmentation of maxillary and mandibular teeth and bone structures for computer-guided therapy.¹³ Dentbird Studio (Imageworks Inc, Seoul, Korea) is capable of 3D landmark detection, automatically identifying a total of 65 landmarks. In the present study, these were detected near their actual positions despite the presence of various devices, such as orthodontic appliances and orthognathic surgical hardware. The PC specifications included a GeForce 2080Ti GPU, 64 GB RAM, and an Intel i7 3.6 GHz CPU. This study aimed to evaluate the clinical validity and accuracy of a deep learning-based CBCT automatic landmarking algorithm in 3D automatic cephalometry and analysis. The authors posited that the values produced by the algorithm would be comparable to those obtained by humans and would promote efficiency by reducing time. Additionally, the authors hypothesized that no errors would be observed in the measured values attributable to hardware and screws after surgery.

Materials and Methods

3D manual landmarking

Dentbird Studio (Imageworks Inc) trained 2 biomedical experts to identify 3D landmarks on 821 CBCT images acquired with an i-CAT 17-19 device (Imaging Sciences International, Hatfield, PA, USA) and 148 CBCT images obtained with various NewTom models (5G, VGi EVO, VGi Mark 3, VGi Mark 4; NewTom, Imola, Italy). This training was conducted under the supervision of a clinician. The experts also recorded the time required to establish the 3D landmarks for each CBCT image. Additionally, they used a stopwatch to measure the duration of the landmarking process for a subset of 30 CBCT images.

3D automatic landmarking

Imageworks Incorporated has developed an automatic 3D landmarking algorithm utilizing a 2-stage coarse-to-fine approach. All CBCT datasets were acquired for diagnostic purposes and exported in Digital Imaging and Communications in Medicine format. The personal data of all patients, including names and registration numbers, were anonymized. The dataset was compiled without regard to sex, age, or race and included perioperative information, with a focus on cases of orthognathic surgery. Consequently, approximately 40% of the CBCT scans contained surgical hardware, such as surgical plates and

Table 1. Anatomical groups and included landmarks

Group	Landmarks
Mid-sagittal Mx. (bone)	A (A-point), ANS (anterior nasal spine), PNS (posterior nasal spine), N (nasion)
Mid-sagittal Mn. (bone)	B (B-point), Pog (pogonion), Gn (gnathion), Me (menton)
Mid-sagittal Mx. (soft tissue)	Sls (soft tissue A-point), Pn (pronasale), Soft N (soft tissue nasion), Sts (stomion superius), Soft Gabella (soft tissue gabella), Ala R (right alar base), Ala L (left alar base)
Mid-sagittal Mn. (soft tissue)	Soft Pog (soft tissue pogonion), Si (mentolabial sulcus), Sti (stomion inferius)
Skull	G (crista galli), Ba (basion), S (sella), Po_R (right porion), Po_L (left porion)
Lateral Mn.	Go_R (right gonion), Go_L (left gonion), M_R (right mental foramen), M_L (left mental foramen), MF_R (right mandibular foramen), MF_L (left mandibular foramen)
Tooth crown (Mx.)	R U1CP (center of right maxillary incisor crown), L U1CP (center of left maxillary incisor crown), R U3CP (tip of right maxillary canine crown), L U3CP (tip of left maxillary canine crown), R U6CP (tip of mesiobuccal cusp of right maxillary first molar crown), L U6CP (tip of mesiobuccal cusp of left maxillary first molar crown)
Tooth roots (Mx.)	R U1RP (root of right maxillary incisor), L U1RP (root of left maxillary incisor), R U3RP (root of right maxillary canine), L U3RP (root of left maxillary canine), R U6RP (mesiobuccal root of right maxillary first molar), L U6RP (mesiobuccal root of left maxillary first molar)
Tooth crown (Mn.)	R L1CP (center of right mandibular incisor crown), L L1CP (center of left mandibular incisor crown), R L3CP (tip of right mandibular canine crown), L L3CP (tip of left mandibular canine crown), R L6CP (tip of mesiobuccal cusp of right mandibular first molar crown), L L6CP (tip of mesiobuccal cusp of left mandibular first molar crown)
Tooth roots (Mn.)	R L1RP (root of right mandibular incisor), L L1RP (root of left mandibular incisor), R L3RP (root of right mandibular canine), L L3RP (root of left mandibular canine), R L6RP (mesiobuccal root of right mandibular first molar), L L6RP (mesiobuccal root of left mandibular first molar)
Bone around orbit	ZyFr_R (right zygomaticofrontal suture), ZyFr_L (left zygomaticofrontal suture), RO_R (right roof of orbit), Or_R (right orbitale), RO_L (left roof of orbit), Or_L (left orbitale)
Condyle	Cl_R (right condylus lateralis), Cm_R (right condylus medialis), Co_R (right condylion), Cl_L (left condylus lateralis), Cm_L (left condylus medialis), Co_L (left condylion)

Mx: maxilla, Mn: mandible

mini-screws.

Sixty-five 3D landmarks in 12 anatomical groups, including the bone, skin, dental crown, tooth root, neural canal (center or opening), and sella, were cataloged (Table 1). The time required to measure these 65 landmarks was 10.9 seconds per volume, with point-to-point errors of 1.7 ± 0.1 mm (99% confidence interval). The threshold for a clinically acceptable successful detection rate (SDR) was set at 3 mm, with SDRs of 88.16% at this level and 94.35% at 4 mm.¹⁴ All 65 landmarks were detected near their true positions, even in the presence of various types of orthognathic surgical hardware.

3D landmark measurement

A total of 80 CBCT scans from Chung-Ang University Hospital in Seoul, Korea (Institutional Review Board number: 1922-007-362), were categorized into a non-sur-

gical group (39 cases) and a surgical group (41 cases). The latter was further subdivided into 9 cases without hardware and 32 cases with hardware. All CBCT scans were anonymized and assigned new serial numbers for the study.

Based on the 65 landmarks, 53 measurements (27 lengths, 21 angles, and 5 ratios) were taken. The classification and measurement values were based on the methodology outlined by Gupta et al.¹¹

The length measurements were categorized into 3 groups: 1) bilateral, obtained from 2 symmetrical landmarks in the parasagittal plane; 2) midsagittal, derived from 2 landmarks in the midsagittal plane; and 3) midsagittal to bilateral, acquired using 3 landmarks—1 in the central sagittal plane and 2 symmetrically located in the parasagittal plane.

The angles were classified into 3 types: midsagittal (calculated using 3 landmarks within the midsagittal plane);

Table 2. Components of \bar{u} and \bar{v} for angular measurement parameters

Angular measurement parameters	Components of \bar{u}	Components of \bar{v}
S-N-A	S, N	N, A
S-N-B	S, N	N, B
A-N-B	A, N	N, B
U1L to ANS-PNS	L U1CP, L U1RP	ANS, PNS
U1R to ANS-PNS	R U1CP, R U1RP	ANS, PNS
U1L-SN	L U1CP, L U1RP	S, N
U1R-SN	R U1CP, R U1RP	S, N
N-GoL-Me	N, Go_L	Go_L, Me
N-GoR-Me	N, Go_R	Go_R, Me
CoL-GoL-Me	Co_L, Go_L	Go_L, Me
CoR-GoR-Me	Co_R, Go_R	Go_R, Me
U1L to L1L (interincisal angle L)	L U1CP, L U1RP	L L1CP, L L1RP
U1R to L1R (interincisal angle R)	R U1CP, R U1RP	R L1CP, R L1RP
L1L to Me-GoL	L L1CP, L L1RP	Me, Go_L
L1R to Me-GoR	R L1CP, R L1RP	Me, Go_R
A-B X N-Pog	A, B	N, Pog
S-N X GoL-Gn	S, N	Go_L, Gn
S-N X GoR-Gn	S, N	Go_R, Gn
ANS-PNS X S-N	ANS, PNS	N, S
PoL-OrL X GoL-Me	Po_L, Or_L	Go_L, Me
PoR-OrR X GoR-Me	Po_R, Or_R	Go_R, Me

midsagittal to bilateral (determined by 4 landmarks, 2 in the midsagittal plane and 2 symmetrically positioned in the parasagittal plane); and planar (either four landmarks or two landmarks and one horizontal plane).

Measurements were obtained through 3D vector calculation. Length was determined by multiplying the 3D voxel index of a landmark by the spacing values along the x, y, and z axes, followed by application of the 3D Euclidean distance formula. For angles, the angle between the vectors \bar{u} and \bar{v} , represented by the pair of landmarks, was calculated (Table 2). This angle was derived using the second law of cosines. When determining the smaller of the angles formed by the 2 vectors, the direction of the vectors was verified and factored into the calculation. In mathematics, the angle between 2 vectors is a value between 0 and 180 degrees. The angle between 2 vectors \bar{u} and \bar{v} can be calculated using the dot product and inverse trigonometric functions.

The following formulas were used to calculate the distances and angles in P(x_1, y_1, z_1), Q(x_2, y_2, z_2), R($x_3, y_3,$

z_3), and S(x_4, y_4, z_4):

$$D_{PQ} = \sqrt{(X_1 - X_2)^2 + (Y_1 - Y_2)^2 + (Z_1 - Z_2)^2}$$

where $X_i = x_i \times (\text{spacing of } x\text{-axis})$,

$Y_i = y_i \times (\text{spacing of } y\text{-axis})$, and

$Z_i = z_i \times (\text{spacing of } z\text{-axis})$

$$\theta_{PQRS} = \cos^{-1} \frac{\bar{u}\bar{v}}{|\bar{u}||\bar{v}|}$$

where $\bar{u} = (x_2 - x_1)\hat{i} + (y_2 - y_1)\hat{j} + (z_2 - z_1)\hat{k}$ and

$\bar{v} = (x_4 - x_3)\hat{i} + (y_4 - y_3)\hat{j} + (z_4 - z_3)\hat{k}$

Statistical analysis

The means and standard deviations (SDs) of the measurements, obtained using landmarks identified by humans or artificial intelligence (AI), were determined. The means, medians, and SDs of the errors between the 2 measured values were also calculated. For the 80 CBCT scans, the Shapiro-Wilk test was applied to assess the normality of the measurements based on human- and AI-identified landmarks. Subsequently, an unpaired *t*-test was used to evaluate whether a significant difference existed between the 2 groups. Additionally, a Bland-Altman plot was employed to visually represent the differences between groups.

An unpaired *t*-test was used to compare the nonsurgical, surgical, hardware present, and hardware absent groups. Differences between the groups were visually expressed using violin plots. Statistical significance was established at $P < 0.05$. Shapiro-Wilk and unpaired *t*-tests were conducted using SPSS GradPacks Statistics 28 (IBM Corp., Armonk, NY, USA). Bland-Altman and violin plots were created using Microsoft Excel (Version 2208 Build 16.0.15601.20148, 64-bit; Microsoft, Redmond, WA, USA).

Results

Comparison of measurement values obtained by human and AI methods

The means and SDs of the measurements, as well as the means, medians, and SDs of the differences between the 2 methods, were calculated (Tables 3-5). The unpaired *t*-test revealed statistically significant differences for 6 of the 53 measured values when comparing landmarks detected by humans and AI (R U6CP-L U6CP, R L3CP-L L3CP, S-N, Or_R-R U3CP, L1L to Me-Go, and GoR-Gn/S-N; $P < 0.05$). The measurements with the highest mean error values were CoL-CoR (3.700 mm) for length, U1R to L1R (3.587°) and U1L to L1L (3.169°) for angle, and

Table 3. Comparison of linear cephalometric measurements between manual and artificial intelligence methods (unit: mm)

Linear measurement parameters	Manual	Artificial intelligence	Error
Bilateral measurement			
ZyFr_R-ZyFr_L	102.09 ± 6.08	101.83 ± 5.15	1.13 ± 1.15
GoL-GoR	92.21 ± 7.73	92.38 ± 7.38	0.94 ± 0.70
CoL-CoR	103.68 ± 6.29	103.24 ± 5.01	3.70 ± 2.50
OrL-OrR	59.63 ± 3.39	59.33 ± 3.12	1.93 ± 1.54
R U3CP-L U3CP	35.09 ± 2.64	34.94 ± 2.08	1.35 ± 1.63
R U6CP-L U6CP	52.66 ± 3.47	52.14 ± 2.79	1.70 ± 1.34*
R L3CP-L L3CP	27.09 ± 1.87	27.78 ± 1.66	1.35 ± 1.01*
R L6CP-L L6CP	47.38 ± 2.99	47.64 ± 2.43	1.63 ± 1.10
Midsagittal measurement			
N-Me	123.15 ± 7.69	122.97 ± 7.36	1.25 ± 1.13
N-ANS	54.39 ± 3.29	54.29 ± 3.00	1.33 ± 1.01
ANS-Me	69.73 ± 6.78	69.62 ± 6.52	1.02 ± 0.94
ANS-PNS	46.92 ± 4.18	47.24 ± 3.44	1.38 ± 1.30
S-N	63.98 ± 3.73	63.60 ± 3.68	0.85 ± 0.74*
Me-L L1CP	42.16 ± 3.53	42.25 ± 3.46	0.68 ± 0.88
Me-R L1CP	42.14 ± 3.50	42.18 ± 3.44	0.62 ± 0.71
Midsagittal to bilateral measurement			
GoL-Pg	88.07 ± 5.73	87.84 ± 5.37	1.69 ± 1.21
GoR-Pg	87.61 ± 5.86	87.61 ± 5.74	1.79 ± 1.38
GoL-N	119.18 ± 7.59	118.81 ± 7.22	1.29 ± 1.16
GoR-N	119.45 ± 7.43	119.01 ± 7.08	1.49 ± 1.40
CoL-GoL	57.1 ± 6.14	56.97 ± 5.86	1.98 ± 1.57
CoR-GoR	57.58 ± 6.25	57.48 ± 5.46	2.12 ± 1.85
CoL-Pg	125.29 ± 7.32	125.28 ± 7.12	1.01 ± 0.80
CoR-Pg	125.43 ± 7.66	125.49 ± 7.67	1.14 ± 1.03
Or_L-L U3CP	55.93 ± 4.28	55.70 ± 4.05	0.76 ± 0.77
Or_R-R U3CP	55.55 ± 4.35	55.33 ± 4.11	0.73 ± 0.62*
Or_L-L U6CP	51.33 ± 4.24	51.38 ± 4.02	0.72 ± 0.57
Or_R-R U6CP	51.51 ± 4.20	51.52 ± 3.96	0.72 ± 0.60

* $P < 0.05$

N-Me/N-ANS (0.043) for ratio.

The limits of agreement and the width obtained in the Bland-Altman analysis are presented in Table 6. The measurements with the largest widths were CoR-CoL (17.49 mm) for length, U1R to L1R (20.06°) for angle, and N-Me/N-ANS (0.22) for ratio.

Comparison of measured values by surgical history and hardware presence

The 80 CBCT scans included a non-surgical group (39 cases) and a surgical group (41 cases), with the latter including 9 cases without hardware and 32 cases with hardware. For each group, measurements derived from human-identified and AI-detected landmarks were com-

pared for the 27 linear parameters using an unpaired *t*-test (Fig. 1). In the linear cephalometric measurements, the non-surgical group displayed a mean error of 1.32 mm and an SD of 1.30 mm. The surgical group without hardware had a mean of 1.10 mm (SD, 1.08 mm), while the group with hardware had a mean of 1.45 mm (SD, 1.49 mm). The lowest error was observed in the surgical group without hardware. No significant difference in measurement agreement was observed between the non-surgical and surgical groups ($P < 0.05$). However, a significant difference was noted between surgical subgroups based on the presence or absence of hardware ($P < 0.05$). These findings were graphically represented using a violin plot (Fig. 2).

Table 4. Comparison of angular cephalometric measurements between manual and artificial intelligence methods (unit: °)

Angular measurement parameters	Manual	Artificial intelligence	Error
Midsagittal measurement			
S-N-A	82.16 ± 3.92	82.27 ± 3.33	1.41 ± 1.22
S-N-B	80.18 ± 4.77	80.24 ± 4.39	1.25 ± 1.15
A-N-B	3.74 ± 2.15	3.64 ± 2.15	0.72 ± 0.56
U1L to ANS-PNS	64.65 ± 7.7	65.13 ± 7.1	2.65 ± 2.1
U1R to ANS-PNS	65.19 ± 7.92	65.29 ± 7.02	2.57 ± 2.4
U1L-SN	104.46 ± 8.15	103.89 ± 7.1	2.37 ± 1.91
U1R-SN	103.94 ± 8.21	103.78 ± 7.03	2.57 ± 1.97
Midsagittal to bilateral measurement			
N-GoL-Me	71.74 ± 4.5	71.8 ± 4.42	0.64 ± 0.61
N-GoR-Me	71.71 ± 4.5	71.8 ± 4.36	0.64 ± 0.69
CoL-GoL-Me	121.28 ± 6.65	121.38 ± 6.41	1.45 ± 1.06
CoR-GoR-Me	121.46 ± 6.16	121.49 ± 6.05	1.38 ± 1.26
U1L to L1L	130.88 ± 10.37	130.62 ± 8.9	3.17 ± 2.91
U1R to L1R	131.18 ± 10.87	131.16 ± 9.37	3.59 ± 3.63
L1L to Me-GoL	96.44 ± 7.77	95.64 ± 6.96	2.74 ± 2.27*
L1R to Me-GoR	94.8 ± 7.4	95.22 ± 7.22	2.72 ± 2.53
Planar measurement			
A-B X N-Pog	5.14 ± 2.77	5.21 ± 3.18	1.16 ± 0.92
S-N X GoL-Gn	47.59 ± 5.23	47.25 ± 5.02	1.55 ± 1.49
S-N X GoR-Gn	46.83 ± 5.25	47.06 ± 4.79	1.5 ± 1.07
ANS-PNS X S-N	168.69 ± 4.57	168.88 ± 4.02	1.61 ± 1.46
PoL-OrLxGoL-Me	30.91 ± 5.78	30.85 ± 5.61	0.94 ± 0.78
PoR-OrRxGoR-Me	30.32 ± 5.35	30.35 ± 5.23	1.12 ± 0.89

**P* < 0.05

Table 5. Comparison of ratios between manual and artificial intelligence methods

Ratio parameters	Manual	Artificial intelligence	Error
N-Me/N-ANS	2.27 ± 0.13	2.27 ± 0.13	0.04 ± 0.04
S-GoL/N-Me	0.73 ± 0.05	0.72 ± 0.05	0.01 ± 0.01
S-GoR/N-Me	0.73 ± 0.05	0.73 ± 0.05	0.02 ± 0.02
GoL-Gn/S-N	1.38 ± 0.09	1.39 ± 0.08	0.04 ± 0.03
GoR-Gn/S-N	1.38 ± 0.09	1.39 ± 0.08	0.03 ± 0.03*

**P* < 0.05

Similarly, for the 21 angle items, the measurements derived from human- and AI-identified landmarks were compared using an unpaired *t*-test (Fig. 3). In the angular cephalometric measurements, the mean error for the non-surgical group was 1.81° (SD, 2.05°). The surgical group without hardware had a mean error of 1.60° (SD, 1.46°), while the surgical group with hardware had a mean error of 1.83° (SD, 1.94°). The surgical group without hardware exhibited the lowest error. No significant dif-

ference was noted between the non-surgical and surgical groups (*P* < 0.05), nor was a significant difference present between surgical subgroups based on the presence or absence of hardware (*P* < 0.05). These findings were graphically represented using a violin plot (Fig. 4).

Finally, measurement values for the 5 ratio items were compared in a similar fashion using an unpaired *t*-test (Fig. 5). For the non-surgical group, the mean error between human- and AI-landmarked measurements was 0.0285 (SD,

Table 6. Bland-Altman analysis of the difference between manual and artificial intelligence-based cephalometric measurements

	95% limit of agreement			95% limit of agreement			
	Upper limit	Lower limit	Width	Upper limit	Lower limit	Width	
	Linear measurement			Angular measurement			
Bilateral measurement				Midsagittal measurement			
ZyFr_R-ZyFr_L	3.38	-2.87	6.24	S-N-A	3.54	-3.77	7.31
GoR-GoL	2.11	-2.44	4.55	S-N-B	3.28	-3.41	6.69
CoR-CoL	9.19	-8.3	17.49	A-N-B	1.88	-1.69	3.57
OrR-OrL	5.12	-4.52	9.64	U1L to ANS-PNS	6.1	-7.06	13.17
R U3CP-L U3CP	4.29	-3.99	8.28	U1R to ANS-PNS	6.8	-7.01	13.81
R U6CP-L U6CP	4.66	-3.61	8.27	U1L-SN	6.45	-5.31	11.75
R L3CP-L L3CP	2.33	-3.72	6.05	U1R-SN	6.53	-6.22	12.74
R L6CP-L L6CP	3.59	-4.1	7.69	Midsagittal to bilateral measurement			
Midsagittal measurement				N-GoL-Me	1.66	-1.8	3.46
N-Me	3.48	-3.12	6.59	N-GoR-Me	1.74	-1.92	3.66
N-ANS	3.38	-3.17	6.55	CoL-GoL-Me	3.43	-3.63	7.06
ANS-Me	2.84	-2.61	5.45	CoR-GoR-Me	3.64	-3.7	7.34
ANS-PNS	3.37	-3.99	7.36	U1L to L1L	8.71	-8.19	16.9
S-N	2.47	-1.71	4.18	U1R to L1R	10.04	-10.01	20.06
Me-L L1CP	2.08	-2.27	4.35	L1L to Me-GoL	7.62	-6.03	13.64
Me-R L1CP	1.8	-1.9	3.7	L1R to Me-GoR	6.84	-7.66	14.5
Midsagittal to bilateral measurement				Midsagittal to bilateral measurement			
GoL-Pg	4.29	-3.82	8.11	A-B X N-Pog	2.84	-2.98	5.82
GoR-Pg	4.45	-4.44	8.9	S-N X GoL-Gn	4.51	-3.82	8.33
GoL-N	3.69	-2.96	6.65	S-N X GoR-Gn	3.37	-3.83	7.2
GoR-N	4.36	-3.47	7.83	ANS-PNS X S-N	4.06	-4.43	8.49
CoL-GoL	5.1	-4.83	9.93	PoL-OrLxGoL-Me	2.46	-2.34	4.81
CoR-GoR	5.63	-5.42	11.05	PoR-OrRxGoR-Me	2.78	-2.84	5.62
CoL-Me	2.53	-2.52	5.05	Ratio parameters			
CoR-Me	2.96	-3.09	6.05	N-Me/N-ANS	0.11	-0.11	0.22
Or_L-L U3CP	2.31	-1.85	4.16	S-GoL/N-Me	0.04	-0.04	0.07
Or_R-R U3CP	2.05	-1.61	3.66	S-GoR/N-Me	0.04	-0.04	0.09
Or_L-L U6CP	1.76	-1.85	3.61	GoL-Gn/S-N	0.08	-0.1	0.18
Or_R-R U6CP	1.81	-1.85	3.66	GoR-Gn/S-N	0.07	-0.09	0.09

0.0275); for the surgical group without hardware, the mean was 0.0188 (SD, 0.0158); and for the surgical group with hardware, the mean was 0.0313 (SD, 0.0297). The surgical group without hardware exhibited the lowest error. No significant difference was noted between the non-surgical and surgical groups ($P > 0.05$). However, a significant difference was noted between surgical subgroups based on the presence or absence of hardware ($P < 0.05$). These findings were graphically represented using a violin plot (Fig. 6).

Discussion

Among the 6 items displaying significant differences, 2 measurements included the left gonion (GoL) or the right gonion (GoR): L1L to Me-GoL and GoR-Gn/S-N ($P < 0.05$). In a previous study by these authors, the SDRs of these landmarks were particularly low. All detection methods from the “2014 Automatic Cephalometric X-Ray Landmark Detection: a grand challenge”, conducted by the Institute of Electrical and Electronics Engineers International Symposium on Biomedical Imaging, misrepresented




Fig. 1. Comparison of linear cephalometric measurements between manual and artificial intelligence methods by patient group. Black error bars represent the 95% confidence standard deviation range for each item value.




Fig. 2. Violin plots representing the difference in linear measurements between the manual and artificial intelligence methods. The thick vertical bar in the violin plot represents the interquartile range, while the thin vertical line indicates the 95% confidence interval; the extremes of this thin line denote the maximum and minimum values. The central white dot signifies the median. The width of a violin plot reflects the density of the data, with wider sections indicating a higher frequency of values and narrower sections representing a lower frequency. The difference values are distributed around the median for all 3 groups: non-surgical, surgical without hardware, and surgical with hardware. A. Violin plots of linear measurements for the 3 groups. B. Violin plots of linear measurements for the non-surgical and surgical groups. C. Violin plots of linear measurements for the surgical group, comparing cases with and without hardware.

the gonion landmark. This resulted in a minimum error greater than 4 mm from the ground truth point.¹⁵ This discrepancy suggests that either the dataset failed to capture the high variability around these landmarks, or errors were present during manual annotation.

Furthermore, of the 6 significant items, 2 measurements involved the sella: S-N and GoR-Gn/S-N ($P < 0.05$). The sella is a fiducial point located at the center of a cavity that, by definition, is a cephalometric landmark easily detectable on 2D head radiographs. However, it is difficult to identify on 3D CBCT because the skull structure does not create 3D contours. Makram et al. proposed a system that automatically localizes 20 three-dimensional hard tissue cephalometric landmarks using Reeb graphs. In their study, the mean error of the sella was notably high, at 2.6 mm.¹⁶ Given the challenges associated with the sella, various methods have been attempted for landmark detection. Montúfar et al. employed a technique involving the circle adjustment of the sub-volume slice of the sella using Hough transformation to generate an anatomical geometric contour of the sella.¹⁷

Four of the 6 items - R U6CP-L U6CP, R L3CP-L L3CP, Or_R-R U3CP, and L1L to Me-GoL ($P < 0.05$) - included landmarks related to the teeth. The identification of landmarks associated with teeth can be affected by the surrounding anatomical structures, leading to potential errors even for clinicians. This is particularly true for the mandibular incisors, which are often difficult to discern due to their typical overlap with the maxillary incisors.¹⁰

In this study, 80 CBCT scans were analyzed. Using unpaired *t*-tests, comparisons were made between measurement values based on manual and AI landmarking.




Fig. 3. Comparison of angular cephalometric measurements between manual and artificial intelligence methods by patient group.




Fig. 4. Violin plots representing the difference in angular measurements between the manual and artificial intelligence methods. The difference values are distributed around the median for all 3 groups: non-surgical, surgical without hardware, and surgical with hardware. A. Violin plots of angular measurements for the 3 groups. B. Violin plots of angular measurements for the non-surgical and surgical groups. C. Violin plots of angular measurements for the surgical group, comparing cases with and without hardware.

Of 53 measurements, statistically significant differences were observed for 4 lengths, 1 angle, and 1 ratio. When

assessing the errors in assessments of length, angle, and ratio based on the designated measurement points on a 3-dimensional structure, the greatest error was found for length. In contrast, even errors at the measurement points had minimal impact on angles and ratios. Given that measured angles and ratios, more so than lengths, are valuable in planning orthodontic treatment or orthognathic surgery, the findings of this study are promising for clinical application.

The 80 CBCT scans were categorized into a non-surgical group (39 cases) and a surgical group (41 cases), with the latter including 9 cases without hardware and 32 cases with hardware. When comparing the manual and AI-based measurements for the 5 ratio items in each group, the cohort in which hardware was removed postoperatively exhibited the lowest measurement error across length, angle, and ratio values. No significant differences were detected in any of the measurement groups when comparing non-surgical and surgical data. In surgical group, however, significant differences in measurement errors for length and ratio were observed depending on whether hardware was present (Figs. 1-6).

Noise in CBCT images, along with metal artifacts from dental prostheses and implants, complicates the accurate delineation of teeth and bones. Hardware and screws were expected to introduce errors; however, the AI method performed well, regardless of hardware and screw presence. Consequently, this algorithm may serve as a valuable tool for assessing the extent of preoperative to postoperative change.

Minnema et al. developed a deep learning algorithm based on mixed-scale density CNNs for the segmentation




Fig. 5. Comparison of cephalometric measurement ratios between manual and artificial intelligence methods by patient group.

of teeth and bones on CBCT images containing metal structures. The algorithm appeared capable of excluding metal artifacts and accurately segmenting teeth and bone structures. These findings indicate that CNNs can identify voxel-level features in CBCT images that humans cannot distinguish.¹⁸

The algorithm employed in the present study was based on deep learning techniques. Dot et al. compared the results of 11 studies to assess the accuracy and reliability of automatic CBCT cephalometric landmarking; the 2 algorithms that demonstrated the best performance employed deep learning methods.¹⁹ In this study, the deep learning-based algorithm reported an average error of less than 2 mm for all landmarks, comparable to the inter-operator variability observed in manual landmarking.

In this study, landmarks were not manually adjusted after 3D automatic landmarking. If manual adjustments were made to landmarks with a high likelihood of error (such as teeth, sellae, and gonions) following automatic landmarking, the accuracy of the measured values could be further improved. Alternatively, hybrid analysis methods that determine specific landmarks through various approaches, such as the Montúfar sella measurement, can be employed.¹⁷

The time required to manually measure 65 landmarks was approximately 40-60 minutes per CBCT volume, although this time was impossible to fully capture because the workers took intermittent breaks. In contrast, the AI algorithm completed the task in 10.9 seconds, with the following PC specifications: GeForce 2080Ti, 64 GB RAM, and an Intel i7 CPU at 3.6 GHz. Since the landmark-based calculation of measurements is identical in the manual and AI methods, the AI method markedly reduced the time needed to identify a landmark and deter-




Fig. 6. Violin plots representing the difference in ratio measurements between the manual and artificial intelligence methods. Difference values are distributed around the median for all 3 groups: non-surgical, surgical without hardware, and surgical with hardware. A. Violin plots of the difference in ratios for the 3 groups. B. Violin plots of the difference in ratios for the non-surgical and surgical groups. C. Violin plots of the difference in ratios for the surgical group, comparing cases with and without hardware.

mine the measurement value.

The accuracy of measurements obtained with the deep learning-based CBCT automatic landmarking algorithm was comparable to that based on human-identified landmarks. By decreasing the time needed to calculate these measurements, the use of such an algorithm can improve the efficiency of diagnosis and treatment.

In this study, measurements of length demonstrated the lowest accuracy. However, as angles and ratios are more commonly utilized than length in patient diagnosis, the findings confirm that employing measurements derived from AI-based landmarks is suitable for diagnostic purposes.

In the comparison between surgical and non-surgical groups, no significant differences were found in linear measurements, angular measurements, or ratio parameters. Similarly, no significant differences were observed between non-surgical and surgical data for any patient group. Additionally, the presence of skeletal deformity did not impact the accuracy of automatic landmark identi-

fication.

The technology discussed in this report is anticipated to increase clinician efficiency and minimize diagnostic errors. This will facilitate the use of 3D cephalometric analyses for clinicians of all experience levels. Furthermore, the availability of a user-friendly, web-based application for 3D automatic landmarking will broaden access for clinicians. At present, no clear standard exists for 3D cephalometric analysis, largely due to the time and effort involved as well as constraints related to its application in corrective procedures, surgical diagnosis, and treatment planning. The findings of this study may assist clinicians in incorporating 3D cephalometric analysis into their practice, irrespective of their level of experience.

Conflicts of Interest: None

References

1. Gribel BF, Gribel MN, Frazão DC, McNamara JA Jr, Manzi FR. Accuracy and reliability of craniometric measurements on lateral cephalometry and 3D measurements on CBCT scans. *Angle Orthod* 2011; 81: 26-35.
2. Lee SH, Kil TJ, Park KR, Kim BC, Kim JG, Piao Z, et al. Three-dimensional architectural and structural analysis - a transition in concept and design from Delaire's cephalometric analysis. *Int J Oral Maxillofac Surg* 2014; 43: 1154-60.
3. Olszewski R, Cosnard G, Macq B, Mahy P, Reyckler H. 3D CT-based cephalometric analysis: 3D cephalometric theoretical concept and software. *Neuroradiology* 2006; 48: 853-62.
4. Mah JK, Huang JC, Choo H. Practical applications of cone-beam computed tomography in orthodontics. *J Am Dent Assoc* 2010; 141 Suppl 3: 7S-13.
5. Lindner C, Wang CW, Huang CT, Li CH, Chang SW, Cootes TF. Fully automatic system for accurate localisation and analysis of cephalometric landmarks in lateral cephalograms. *Sci Rep* 2016; 6: 33581.
6. Vandaele R, Aceto J, Muller M, Peronnet F, Debat V, Wang CW, et al. Landmark detection in 2D bioimages for geometric morphometrics: a multi-resolution tree-based approach. *Sci Rep* 2018; 8: 538.
7. Lagravère MO, Low C, Flores-Mir C, Chung R, Carey JP, Heo G, et al. Intraexaminer and interexaminer reliabilities of landmark identification on digitized lateral cephalograms and formatted 3-dimensional cone-beam computerized tomography images. *Am J Orthod Dentofacial Orthop* 2010; 137: 598-604.
8. Hassan B, Nijkamp P, Verheij H, Tairie J, Vink C, van der Stelt P, et al. Precision of identifying cephalometric landmarks with cone beam computed tomography in vivo. *Eur J Orthod* 2013; 35: 38-44.
9. Ghowsi A, Hatcher D, Suh H, Wile D, Castro W, Krueger J, et al. Automated landmark identification on cone-beam computed tomography: accuracy and reliability. *Angle Orthod* 2022; 92: 642-54.
10. Jeon S, Lee KC. Comparison of cephalometric measurements between conventional and automatic cephalometric analysis using convolutional neural network. *Prog Orthod* 2021; 22: 14.
11. Gupta A, Kharbanda OP, Sardana V, Balachandran R, Sardana HK. Accuracy of 3D cephalometric measurements based on an automatic knowledge-based landmark detection algorithm. *Int J Comput Assist Radiol Surg* 2016; 11: 1297-309.
12. Barrett JF, Keat N. Artifacts in CT: recognition and avoidance. *Radiographics* 2004; 24: 1679-91.
13. Hung K, Yeung AW, Tanaka R, Bornstein MM. Current applications, opportunities, and limitations of AI for 3D imaging in dental research and practice. *Int J Environ Res Public Health* 2020; 17: 4424.
14. Torosdagli N, Liberton DK, Verma P, Sincan M, Lee JS, Bagci U. Deep geodesic learning for segmentation and anatomical landmarking. *IEEE Trans Med Imaging* 2019; 38: 919-31.
15. Lee JH, Yu HJ, Kim MJ, Kim JW, Choi J. Automated cephalometric landmark detection with confidence regions using Bayesian convolutional neural networks. *BMC Oral Health* 2020; 20: 270.
16. Makram M, Kamel H. Reeb graph for automatic 3D cephalometry. *Int J Image Process* 2014; 8: 17-29.
17. Montufar J, Romero M, Scougall-Vilchis RJ. Hybrid approach for automatic cephalometric landmark annotation on cone-beam computed tomography volumes. *Am J Orthod Dentofacial Orthop* 2018; 154: 140-50.
18. Minnema J, van Eijnatten M, Hendriksen AA, Liberton N, Pelt DM, Batenburg KJ, et al. Segmentation of dental cone-beam CT scans affected by metal artifacts using a mixed-scale dense convolutional neural network. *Med Phys* 2019; 46: 5027-35.
19. Dot G, Rafflenbeul F, Arbotto M, Gajny L, Rouch P, Schouman T. Accuracy and reliability of automatic three-dimensional cephalometric landmarking. *Int J Oral Maxillofac Surg* 2020; 49: 1367-78.